

# Florida Department of Environmental Protection Safe Drinking Water Program Laboratory Reporting Format

## INSTRUCTIONS

This format is to be used by laboratories for reporting drinking water analyses to the appropriate Department of Environmental Protection (DEP) or Health Department offices (see below). For analysis results to be acceptable for compliance with Chapter 62-550, F.A.C., laboratories performing the analyses must be certified for those analyses in drinking water by the Florida Department of Health (DOH) and report results in accordance with Chapter 62-160, F.A.C. Computer generated or otherwise personalized reports will be accepted as long as they conform to this format, with the results reported using the same groups with the same contaminant IDs, contaminant names, units, column headings, and in the same order as in this format.

The contaminants shown in this example are those named in Rule 62-550, F.A.C. Other contaminants may be reported using this format as long as the contaminant IDs and names are the same as those found in the Safe Drinking Water Act. When applicable, results shall be reported using the data qualifier codes specified in the attached Table 1 of Rule 62-160, F.A.C. Parameters not reported need not be listed. Multiple samples for the same parameter can be reported under the same cover sheet as long as the sample locations are specific. Likewise, composite samples must be so identified so that if samples are from multiple plants and/or systems, credit for compliance will be given to all the plants/systems. Radionuclides samples that have been composited over time (multiple quarters) also have to be clearly identified so that proper credit for compliance will be given.

Send completed reports to the DEP District Office or DOH Health Department office which has jurisdiction over the water system. The 7-character PWS ID must be included on the report.

Northwest DEP District Office.....	160 Governmental Center, Pensacola, FL, 32502-5794.....	850.595-8300
Northeast DEP District Office .....	8800 Baymeadows Way West, Suite 100, Jacksonville, FL, 32256.....	904.256-1700
Central DEP District Office .....	3319 Maguire Blvd., Suite 232, Orlando, FL, 32803-3767.....	407.894-7555
Southeast DEP District Office .....	400 N. Congress Ave., Suite 200, West Palm Beach, FL, 33416.....	561.681-6600
Southeast DEP District Branch Office .....	1801 S.E. Hillmoor Dr., Suite C-204, Port St. Lucie, FL, 34952.....	772.398-2806
South DEP District Office .....	P.O. Box 2549, Ft. Myers, FL, 33902-2549 .....	239.332-6975
Southwest DEP District Office .....	13051 N. Telecom Parkway, Temple Terrace, FL, 33637.....	813.632-7600
Broward County Health Department .....	2421 S.W. 6th Ave., Ft. Lauderdale, FL, 33315.....	954.467-4854
Dade County Health Department .....	1725 N.W. 167th St., Suite 119, Miami, FL, 33056.....	305.623-3552
Hillsborough County Health Department .....	P.O. Box 5135, Tampa, FL, 33675-5135.....	813.307-8059
Lee County Health Department .....	.60 Danley Dr., Unit 1, Ft. Myers, FL, 33907 .....	239.274-2200
Palm Beach County Health Department.....	P.O. Box 29, West Palm Beach, FL, 33402 .....	561.837-5900
Polk County Health Department.....	1290 Golfview Ave., 4 <sup>th</sup> Floor, Bartow, FL, 33830.....	863.519-8330
Sarasota County Health Department .....	1301 Cattleman Road, Sarasota, FL, 34232.....	941.378-6133
Volusia County Health Department.....	Environmental Engineering – Bin #180, Box 9190, Daytona Beach, FL, 32120-9190...	386.736-5436

# Florida Department of Environmental Protection Safe Drinking Water Program Laboratory Reporting Format

## DATA QUALIFIER CODES (From 62-160, Table 1)

The following qualifier codes shall be used by laboratories when reporting data values that either meet the specified descriptions outlined below or do not meet the quality control criteria of the laboratory. They are categorized on these 2 pages as Acceptable, Not Acceptable, or May Be Acceptable for compliance.

The following codes (B,D,E,I,K,L,M,U,V,!) <b>ARE ACCEPTABLE</b> for use with results submitted for compliance with 62-550 and 62-555.	
SYMBOL	MEANING
B	Results based upon colony counts outside the acceptable range. Applies to microbiological tests and specifically to membrane filter colony counts. It is to be used if the colony count is generated from a plate in which the total number of coliform colonies is outside the method indicated ideal range. This code is not to be used if a 100 mL sample has been filtered and the colony count is less than the lower value of the ideal range.
D	Measurement was made in the field (i.e., in situ). This code applies to any value (except field measurements of pH, specific conductance, dissolved oxygen, temperature, total residual chlorine, transparency, turbidity or salinity) that was obtained under field conditions using approved analytical methods. If the parameter code specifies a field measurement (e.g., "Field pH"), this code is not required.
E	Indicates that extra samples were taken at composite stations.
I	The reported value is greater than or equal to the laboratory method detection limit but less than the laboratory practical quantitation limit.
K	Off-scale low. Actual value is known to be less than the value given. This code shall be used if the value is less than the lowest calibration standard and the calibration curve is known to be non-linear; or the value is known to be less than the reported value based on sample size, dilution. Shall not be used to report values that are less than the laboratory practical quantitation limit or laboratory method detection limit.
L	Off-scale high. Actual value is known to be greater than value given. To be used when the concentration of the analyte is above the acceptable level for quantitation (exceeds the linear range or highest calibration standard) and the calibration curve is known to exhibit a negative deflection.
M	When reporting chemical analyses: presence of material is verified but not quantified; the actual value is less than the value given. The reported value shall be the laboratory practical quantitation limit. This code shall be used if the level is too low to permit accurate quantification, but the estimated concentration is greater than or equal to the method detection limit. If the value is less than the method detection limit use "T" below.
U	Indicates that the compound was analyzed for but not detected. This symbol shall be used to indicate that the specified component was not detected. The value associated with the qualifier shall be the laboratory method detection limit. Unless requested by the client, less than the method detection limit values shall not be reported (see "T" below).
V	Indicates that the analyte was detected at or above the method detection limit in both the sample and the associated method blank and the value of 10 times the blank value was equal to or greater than the associated sample value. Note: unless specified by the method, the value in the blank shall not be subtracted from associated samples.
!	Data deviate from historically established concentration ranges.
The following codes (A,F,H,N,O,T,Z,?,*) are <b>NOT ACCEPTABLE</b> for use with results submitted for compliance with 62-550 and 62-555.	
A	Value reported is the arithmetic mean (average) of two or more determinations. This code shall be used if the reported value is the average of results for two or more discrete and separate samples. These samples shall have been processed and analyzed independently. Do not use this code if the data are the result of replicate analysis on the same sample aliquot, extract or digestate.
F	When reporting species: F indicates the female sex.
H	Value based on field kit determination; results may not be accurate. This code shall be used if a field screening test (i.e., field gas chromatograph data, immunoassay, vendor-supplied field kit, etc.) was used to generate the value and the field kit or method has not been recognized by the Department as equivalent to laboratory methods.

# Florida Department of Environmental Protection Safe Drinking Water Program Laboratory Reporting Format

## DATA QUALIFIER CODES (From 62-160, Table 1) CONTINUED

N	Presumptive evidence of presence of material. This qualifier shall be used if the component has been tentatively identified based on mass spectral library search; or there is an indication that the analyte is present, but quality control requirements for confirmation were not met (i.e., presence of analyte was not confirmed by alternative procedures).
O	Sampled, but analysis lost or not performed.
T	Value reported is less than the laboratory method detection limit. The value is reported for informational purposes only and shall not be used in statistical analysis.
Z	Too many colonies were present for accurate counting. Historically, this condition has been reported as “too numerous to count” (TNTC). The “Z” qualifier code shall be reported when the total number of colonies of all types is more than 200 in all dilutions of the sample. When applicable to the observed test results, a numeric value for the colony count for the microorganism tested shall be estimated from the highest dilution factor (smallest sample volume) used for the test and reported with the qualifier code.
?	Data are rejected and should not be used. Some or all of the quality control data for the analyte were outside criteria, and the presence or absence of the analyte cannot be determined from the data.
*	Not reported due to interference.
The following codes (J,Q,R,Y) <b>MAY OR MAY NOT BE ACCEPTABLE</b> for use with results submitted for compliance with 62-550 and 62-555, depending on the parameter(s) and/or the circumstances. Results with these codes will be evaluated on a case by case basis.	
SYMBOL	MEANING
J	Estimated value. A “J” value shall be accompanied by a detailed explanation for designating the value as estimated. Where possible, the lab shall report whether the actual value is estimated to be less than or greater than the reported value. A “J” value shall not be used as a substitute for K, L, M, T, V, or Y, however, if additional reasons exist for identifying the value as an estimate (e.g., matrix spiked failed to meet acceptance criteria), the “J” code may be added to a K, L, M, T, V, or Y. Examples of situations in which a “J” code must be reported include: instances where a quality control item associated with the reported value failed to meet the established quality control criteria (the specific failure must be identified); instances when the sample matrix interfered with the ability to make any accurate determination; instances when data are questionable because of improper laboratory or field protocols (e.g., composite sample was collected instead of a grab sample); instances when the analyte was detected at or above the method detection limit in a blank other than the method blank (such as calibration blank or field-generated blanks and the value of 10 times the blank value was equal to or greater than the associated sample value); or instances when the field or laboratory calibrations or calibration verifications did not meet calibration acceptance criteria.
Q	Sample held beyond the accepted holding time. This code shall be used if the value is derived from a sample that was prepared or analyzed after the approved holding time restrictions for sample preparation or analysis.
R	Significant rain in the past 48 hours. (Significant rain typically involves rain in excess of 1/2 inch within the past 48 hours.) This code shall be used when the rainfall might contribute to a lower than normal value.
Y	The laboratory analysis was from an improperly preserved sample. The data may not be accurate.

# Florida Department of Environmental Protection Safe Drinking Water Program Laboratory Reporting Format

## PUBLIC WATER SYSTEM INFORMATION (to be completed by sampler – please type or print legibly)

System Name: HYDRALIFE WATER SERVICES PWS I.D. #: \_\_\_\_\_

System Type (check one):  Community  Nontransient Noncommunity  Transient Noncommunity

Address: 5041 S STATE ROAD SEVEN #409

City: DAVIE ZIP Code: 33314

Phone # (954) 822-9257 Fax #: \_\_\_\_\_ E-Mail Address: hydralifewaterservices@gmail.com

## SAMPLE INFORMATION (to be completed by sampler)

Sample Number: 41545 Sample Date: March 13, 2024 Sample Time: 12:00 AM PM (Circle One)

Sample Location (be specific): BOTTLING DISTRIBUTION Location Code: \_\_\_\_\_

Disinfectant Residual (Required when reporting results for trihalomethanes and haloacetic acids): 0.0 mg/L Field pH: 9.0

### Sample Type (Check Only One)

### Reason(s) for Sample (Check all that apply)

Distribution

Routine Compliance with 62-550

Replacement (of Invalidated Sample)

Entry Point (to Distribution)

Confirmation of MCL Exceedance\*

Special (not for compliance with 62-550)

Plant Tap (not for compliance with 62-550)

Composite of Multiple Sites\*\*

Clearance (permitting)

Raw (at well or intake)

Other: QUALITY CONTROL

Max Residence Time

Sampling Procedure Used or Other Comments:

Ave Residence Time

Near First Customer

\*See 62-550.500(6) for requirements and restrictions.  
And 62-550.512(3) for nitrate or nitrite exceedances.

\*\*See 62-550.550(4) for requirements and  
attach a results page for each site.

## SAMPLER CERTIFICATION

I, JULIE FURMAN, LAB MANAGER, do HEREBY CERTIFY  
(Print Name) (Print Title)

that the above public water system and sample collection information is complete and correct.

Signature: *Julie Furman*

Date: 4-18-26

Certified Operator #: \_\_\_\_\_ Phone #: \_\_\_\_\_

Sampler's Fax #: \_\_\_\_\_

Sampler's E-mail: \_\_\_\_\_

# Florida Department of Environmental Protection Safe Drinking Water Program Laboratory Reporting Format

**LABORATORY CERTIFICATION INFORMATION** (to be completed by lab – please type or print legibly)

Lab Name: ERS LABORATORY/ AEL Florida subcontracted DOH Certification #: E82535 Certification Exp Date: JUNE 30, 2026

**ATTACH CURRENT DOH ANALYTE SHEET\***

Address: 3860 SW 30<sup>TH</sup> AVE FORT LAUDERDALE FL 33312 Phone #: 954-316-8792

Were any analyses subcontracted?  Yes  No If yes, please provide DOH certification number(s): E82535

**ATTACH DOH ANALYTE SHEET FOR EACH SUBCONTRACTED LAB\***

**ANALYSIS INFORMATION** (to be completed by lab) Date Sample(s) Received: March 13, 2026

PWS ID (From Page 1): \_\_\_\_\_ Sample Number (From Page 1): 41545 Lab Assigned Report # or Job ID: 40401


Group(s) Analyzed & Results attached for compliance with Chapter 62-550, F.A.C. (Check all that apply):

<u>Inorganics</u>	<u>Synthetic Organics</u>	<u>Volatile Organics</u>	<u>Disinfection Byproducts</u>	<u>Radionuclides</u>	<u>Secondaries</u>
<input checked="" type="checkbox"/> All Except Asbestos	<input type="checkbox"/> All 30	<input type="checkbox"/> All 21	<input type="checkbox"/> Trihalomethanes	<input type="checkbox"/> Single Sample	<input checked="" type="checkbox"/> All 14
<input type="checkbox"/> Partial	<input type="checkbox"/> All Except Dioxin	<input type="checkbox"/> Partial	<input type="checkbox"/> Haloacetic Acids	<input type="checkbox"/> Qtrly Composite**	<input type="checkbox"/> Partial
<input type="checkbox"/> Nitrate	<input type="checkbox"/> Partial		<input type="checkbox"/> Chlorite		
<input type="checkbox"/> Nitrite	<input type="checkbox"/> Dioxin Only		<input type="checkbox"/> Bromate		
<input type="checkbox"/> Asbestos					

## LAB CERTIFICATION

I, JULIE FURMAN, LAB MANAGER, do HEREBY CERTIFY  
(Print Name) (Print Title)

that all attached analytical data are correct and unless noted meet all requirements of the National Environmental Laboratory Accreditation Conference (NELAC).

Signature:  Date: 4-18-26

\* Failure to provide a valid and current Florida DOH lab certification number and a current Analyte Sheet for the attached analysis results will result in rejection of the report, possible enforcement against the public water system for failure to sample, and may result in notification of the DOH Bureau of Laboratory Services.

\*\* Please provide radiological sample dates & locations for each quarter.

**CONFIRMATION & NOTIFICATION IS REQUIRED WITHIN 24 HRS FOR NITRATE OR NITRITE MCL EXCEEDANCES  
NON-DETECTS ARE TO BE REPORTED AS THE MDL WITH A "U" QUALIFIER. (Non-detects reported as "BDL" or with a "<" are not acceptable.)**

**COMPLIANCE DETERMINATION** (to be completed by DEP or DOH -- attach notes as necessary)

Sample Collection & Analysis Satisfactory:  Yes  No \_\_\_\_\_ Replacement Sample or Report Requested (circle or highlight group(s) above)

Person Notified: \_\_\_\_\_ Date Notified: \_\_\_\_\_ DEP/DOH Reviewing Official: \_\_\_\_\_

# Florida Department of Environmental Protection Safe Drinking Water Program Laboratory Reporting Format

INORGANIC CONTAMINANTS  
62-550.310(1)

Report Number / Job ID: 41545

PWS ID (From Page 1) \_\_\_\_\_

Contam ID	Contam Name	MCL	Units	Analysis Result	Qualifier*	Analytical Method	Lab MDL	Analysis Date	Analysis Time	DOH Lab Certification #
1040	Nitrate (as N)	10	mg/L	0.050	U	EPA353.2	0.025	03/13/26	14:34	<u>E82535</u>
1041	Nitrite (as N)	1	mg/L	0.025	U	EPA353.2	0.025	03/13/26	14:34	<u>E82535</u>
1005	Arsenic	0.010	mg/L	0.00050	U	EPA200.8	0.00050	03/18/26	09:45	<u>E82535</u>
1010	Barium	2	mg/L	0.00090	U	EPA200.8	0.00090	03/18/26	09:45	<u>E82535</u>
1015	Cadmium	0.005	mg/L	0.00032	U	EPA200.7	0.00032	03/18/26	09:45	<u>E82535</u>
1020	Chromium	0.1	mg/L	0.00071	U	EPA200.7	0.00071	03/18/26	09:45	<u>E82535</u>
1024	Cyanide	0.2	mg/L	0.0040	U	EPA335.4	0.0040	03/18/26	09:45	<u>E82535</u>
1025	Fluoride	4.0	mg/L	0.034	U	EPA300.0	0.034	03/13/26	09:45	<u>E82535</u>
1030	Lead	0.015	mg/L	0.00010	U	EPA200.8	0.00010	03/18/26	09:45	<u>E82535</u>
1035	Mercury	0.002	mg/L	0.000065	U	EPA245.1	0.000065	03/18/26	15:22	<u>E82535</u>
1036	Nickel	0.1	mg/L	0.0023	U	EPA200.7	0.0023	03/19/26	19:07	<u>E82535</u>
1045	Selenium	0.05	mg/L	0.00090	U	EPA200.8	0.00090	03/19/26	19:07	<u>E82535</u>
1052	Sodium	160	mg/L	13.1		EPA200.7	0.27	03/19/26	19:07	<u>E82535</u>
1074	Antimony	0.006	mg/L	0.00070	U	EPA200.8	0.00070	03/19/26	19:07	<u>E82535</u>
1075	Beryllium	0.004	mg/L	0.000010	U	EPA200.8	0.000010	03/19/26	19:07	<u>E82535</u>
1085	Thallium	0.002	mg/L	0.00040	U	EPA200.8	0.00040	03/19/26	19:07	<u>E82535</u>
	Potassium	n/a	mg/L	7		EPA200.7	1	03/19/26	19:07	<u>E82535</u>
	Magnesium	n/a	mg/L	1.2		EPA200.7	0.1	03/19/26	19:07	<u>E82535</u>

\*Results must be reported with appropriate qualifiers in accordance with Florida Administrative Code Rule 62-160, Table 1. Results qualified with A, F, H, N, O, T, Z, ?, \*, are unacceptable for compliance with 62-550. Results qualified with a J, Q, R, or Y must be accompanied by written justification and will be evaluated on a case by case basis. To avoid a monitoring violation, unacceptable results must be replaced with acceptable results from samples collected during the same monitoring period.

# Florida Department of Environmental Protection Safe Drinking Water Program Laboratory Reporting Format

SECONDARY CONTAMINANTS  
62-550.320

Report Number / Job ID: 41545

PWS ID (From Page 1): \_\_\_\_\_

Contam ID	Contam Name	MCL	Units	Analysis Result	Qualifier*	Analytical Method	Lab MDL	Analysis Date	Analysis Time	DOH Lab Certification #
1002	Aluminum	0.2	mg/L	0.052	U	EPA200.7	0.052	03/18/26	09:45	<u>E82535</u>
1017	Chloride	250	mg/L	0.66	U	SM4500Cl-C	0.66	03/13/26	14:34	<u>E82535</u>
1022	Copper	1	mg/L	0.0085	U	EPA200.8	0.0085	03/18/26	09:45	<u>E82535</u>
1025	Fluoride	2.0	mg/L	0.034	U	EPA300.0	0.034	03/18/26	09:45	<u>E82535</u>
1028	Iron	0.3	mg/L	0.018	U	EPA200.7	0.018	03/18/26	09:45	<u>E82535</u>
1032	Manganese	0.05	mg/L	0.0040	U	EPA200.8	0.0040	03/18/26	09:45	<u>E82535</u>
1050	Silver	0.1	mg/L	0.00069	U	EPA200.7	0.00069	03/18/26	09:45	<u>E82535</u>
1055	Sulfate	250	mg/L	2.1		EPA300.0	0.5	03/18/26	09:45	<u>E82535</u>
1095	Zinc	5	mg/L	0.0050	U	EPA200.7	0.0050	03/18/26	09:45	<u>E82535</u>
1905	Color	15	CU	5.0	U	SM2120B	5.0	03/14/26	16:08	<u>E82535</u>
1920	Odor	3	TON	<1	U	SM2150B	1.0	03/14/26	14:12	<u>E82535</u>
1925	pH (field pH from page 1)	6.5 - 8.5		9.0		SM4500H <sup>+</sup> -B	0.01	03/13/26	13:30	<u>E82535</u>
1930	Total Dissolved Solids	500	mg/L	38		SM2540C	5.0	03/18/26	16:51	<u>E82535</u>
2905	Foaming Agents	0.5	mg/L	0.40	U	SM5540C	0.40	03/20/26	13:04	<u>E82535</u>